

AMENDMENTS**In the Specification:**

Please add paragraph [0036.1] after paragraph [0036] as follows:

a1
[0036.1] Fig. 18 shows a catheter assembly having a catheter shaft attached to a fluid delivery attachment at a proximal end and to the self-expanding balloon assembly at a distal end.

In the Claims:

Please substitute the claims as indicated below for the claims of the same number.

Claims 1 and 32 have been amended (as shown later in this Amendment). A clean copy of amended claims 1 and 32 are below.

Sub 3
a2
1. (Amended) An embolizing device for insertion into an aneurysm, comprising:
at least one self-expanding member configured to be sealed within a membrane;
the membrane defining a volume and further defining at least one orifice in a surface of
the membrane.

a3
Sub 3
32. (Amended) The embolizing device of claim 1 wherein the membrane comprises a wall having a thickness of about 0.001 inches.